

## 90-1V-High-voltage tester



### High voltage tester DC 90-1V

- Test voltage 0 .. 6,000 VDC grounded
- Tripping current 0 .. 2.00 mA
- Current-limited according to EN50191
- Manual operation
- Interface operation

#### HV test

- Voltage manually adjustable: erdgebunden
- Test voltage (general): 0 .. 6.000 VDC
- Tripping current: 0.1 .. 2.0 mA
- Current limitation according to EN50191
- Tripping current manually adjustable
- Contact monitoring by reading the test voltage

#### Display / Operation

- digital display 3 1/2 -digit

#### Interfaces

- digital-/ analog interface
- Analog interface

#### Equipment data

- Mains voltage: 230 +/- 10%
- Safety class: I
- Slide in technique: 19"
- Height : 3 HU

#### Extensions / Options

- Factory calibration certificate (optional)

## 90-1V-High-voltage tester

---

### **Order data**

- Order description: High-voltage tester
- Order number: 90-1V.3